

YSI & Cable Serial Numbers:

	Reading	Target	Acceptable	Pass
Temp	15.6	Ref: 15.5	$\pm 1^{\circ}\text{C}$	✓
pH7mv	-27.9	0.0	0 ± 50	✓
pH4mv	145.8	177	177 ± 50	✓
pH Slope	173.7	177	162 - 180	✓
Cond. Cell Constant	4.9	5	4.6 - 5.4	✓
Redox Offset	16.4	0.0	± 50.0	✓
DO Gain	Pass or fail determined by the meter			✓

All parameters were within acceptable range on the day of despatch; however we do recommend that the instrument is calibrated daily to ensure accurate readings.

* Calibrated to manufacturers standards

Date: 27.03.17