

YSIPRO Calibration

 YSI & Cable Serial Numbers: 15F100788

	Reading	Target	Acceptable	Pass
Temp	16.7	Ref: 16.9	$\pm 1^{\circ}\text{C}$	✓
pH7mv	-0.1	0.0	0 ± 50	✓
pH4mv	172.3	177	177 ± 50	✓
pH Slope	172.4	177	162 - 180	✓
Cond. Cell Constant	4.9	5	4.6 - 5.4	✓
Redox Offset	7.5	0.0	± 40.0	✓
DO Gain	Pass or fail determined by the meter			✓

All parameters were within acceptable range on the day of despatch; however we do recommend that the instrument is calibrated daily to ensure accurate readings.

 Signed: 

Date: 27.01.16

 Name: 